



3700 Sagamore Parkway N | Lafayette, IN 47904 USA
Tel: (765) 423-1505 | Fax: (765) 423-4111
info@lafayetteinstrument.com | www.lafayetteinstrument.com

Calibration and Functionality Checks for Lafayette Instrument Polygraph Systems

All Lafayette Instrument computerized polygraph systems, including the LX4000 and LX5000 models, are delivered from the factory with all necessary calibration procedures completed. Users of Lafayette Polygraph systems are not required to perform any routine calibration or functionality checks to ensure the proper functioning and accuracy of the equipment. Users may choose to perform periodic functionality checks to verify proper operation in accordance with their local governing guidelines. In keeping with ASTM standards and best practices as defined by the American Polygraph Association, Lafayette Instrument Company recommends performing functionality checks when the examiner suspects a functionality problem.

A handwritten signature in black ink, appearing to read 'Brent E. Smitley'.

Brent Smitley
Engineering Manager
Lafayette Instrument Company